

803



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Mar 16 2006 INITIALS: DEC OWNER: WSP
DISTRICT: 8 ADDRESS: _____
LOCATION: 9th St Shop COUNTY: _____
MAKE: Hanni TYPE: port. wheel weight ACCURACY CLASS: III
MODEL: 101 ID. NO: 21829 CAPACITY: 20,000 lbs SECTIONS: 1
MINIMUM GRADUATIONS: 50 lb REG. ELEMENT: dia PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Center</u>	<u>5000</u>	<u>5000</u>	<u>0</u>				
<u>"</u>	<u>10000</u>	<u>10000</u>	<u>0</u>				
<u>"</u>	<u>15000</u>	<u>15000</u>	<u>0</u>				
<u>Center</u>	<u>20000</u>	<u>20000</u>	<u>0</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify, under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by:

David Cronin

Date:

3/28/06

State of Washington

County of Thurston

Signed and sworn to (or affirmed) before
me on March 28, 2006 by

Michael J. Evans
(Signature of notary public)

Printed Name of Notary

My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS